

# CHECKLIST FOR CERTIFICATE OF CALIBRATION/VALIDATION/TESTING REPORTS

## A

If Certificate Identifies an Accredited Laboratory:



- ILAC/MRA Signatory body accredited Laboratory**

The Following Table lists the accredited laboratories

| A2LA | L-A-B | ACLASS | IAS | PJLA | NVLAP |
|------|-------|--------|-----|------|-------|
|      |       |        |     |      |       |

**AND**

- Name of Device (Optional)**
  - Model Number**
  - Serial Number**
  - Date of Calibration (Report or Issue Date)**
  - Measurement results indicate unit passed test and the documented uncertainty is within suitable limits (recommended uncertainty = +/- 1F (0.5C))**
- 

## B

If Certificate Does Not Identify an Accredited Laboratory:

- Name of Device (Optional)**
- Model Number**
- Serial Number**
- Date of calibration testing (Report or Issue Date)**
- Measurement results indicate unit passed test and the documented uncertainty is within suitable limits (recommended uncertainty = +/- 1F (0.5C))**
- Statement that calibration testing conforms to ISO 17025***

# Certificate Of Calibration

Digital Thermometer W Thermistor Probe  
Report No. 0926



Calibration Laboratory 23

Customer: TAGE HOSPITAL  
185 GRAFT RD  
TOWNS, VA 00216

Date Received: 09/26/2012  
Calibration Date: 09/26/2012

Make: TROL COP  
Model: 4ICC with P10 PROBE  
Serial #: 8042  
/Range: -200 TO 800 °C IN 0.01 °C DIVISIONS  
Accuracy/Tolerance: +/- 0.1 % + 0.2 °C BELOW 200 °C  
Item Received: IN TOLERANCE  
Calibration Location: SCH Temperature Laboratory

Customer Specified Due Date: 09/2013  
PO#: 011513  
Contact: JAY BELCHER  
Temperature: 21.6 TO 21.8 °C / RH% 47 TO 47  
CONDITION RECEIVED: IN SPEC  
Item Returned: IN TOLERANCE  
Equipment Location: LAB

Notes: CALIBRATED AT CUSTOMERS SPECIFIED POINTS OF USE ONLY!

| Nominal | Actual (STD) | Measured (UUT) | Deviation (UUT) | Units | Tolerance (±) | Uncertainty (±) | Pass/Fail |
|---------|--------------|----------------|-----------------|-------|---------------|-----------------|-----------|
| 0       | 0.028        | 0.08           | 0.05            | °C    | 0.20          | 0.09            | PASS      |
| 20      | 20.017       | 20.15          | 0.13            | °C    | 0.22          | 0.09            | PASS      |
| 35      | 35.003       | 35.20          | 0.20            | °C    | 0.24          | 0.09            | PASS      |

Deviation rounded to the readability of UUT

The measurement traceability and calibration process used for conformance verification of the above instrument meets or exceeds the requirements of 17025:2005. The reported uncertainties reflect those of type B (Systematic errors associated with the standards and the procedure used), and type A (Random errors of the process). The type A and type B uncertainties were calculated in accordance with NIST technical Note 1297 using the RSS method and are reported at the coverage factor k=2 to approximate a confidence level of 95%. The due date as it appears on this report does not imply that the instrument will maintain its accuracy for any given length of time unless supported with further documentation (E.g. statistical etc.) which affirms such stability and is the responsibility of the end user. Many factors may contribute to instrument in-accuracy over time such as drift, environment, transportation, frequency of use etc. The reported results reflect readings obtained at the time of test only. The reported uncertainties reflect those associated with the calibration process itself and not the instrument under test. If the UUT is a digital electronic measurement instrument add 0.6 of the least significant digit to the above stated uncertainty. The instrument is considered to be in-tolerance based on the observed results (Deviation or departure from nominal value) falling anywhere within its specified tolerance limits without consideration of applied uncertainty. This document shall not be reproduced except in full without the written approval of Q.C. Services, Inc.  
Procedure Used QCS 3015 (ORIG) (QCSTD 030106-3)

**TRACEABLE STANDARDS USED:**

|                  |             |                  |   |
|------------------|-------------|------------------|---|
| Fluke 1522       | S/N: A6C265 | Cal Due: 10/2012 |   |
| ERTCO-EUTECHNICS | S/N: 304526 | Cal Due: 01/2013 | X |
| HART SCI 1502    | S/N 8B552   | Cal Due: 04/2013 | X |

Certified by: Howard Richard

Date: 09/26/2012

Approved By: 

Title: Metrologist

Date: 09/26/2012

**Example**

**1**

# Certificate Of Calibration

Digital Thermometer W Thermistor Probe

Report No. 0926



Customer: TAGE HOSPITAL  
185 GRAFT RD  
TOWNS, VA 00216

Date Received: 09/26/2012  
Calibration Date: 09/26/2012

Make: TROL COP

Customer Specified Due Date: 09/2013

Model: 41CC with P10 PROBE

PO#: 011513

Serial #: 8042

Contact: JAY BELCHER

/Range: -200 TO 800 °C IN 0.01 °C DIVISIONS

Temperature: 21.6 TO 21.8 °C / RH% 47 TO 47

Accuracy/Tolerance: +/- 0.1 % + 0.2 °C BELOW 200 °C

CONDITION RECEIVED: IN SPEC

Item Received: IN TOLERANCE

Item Returned: IN TOLERANCE

Calibration Location: SCH Temperature Laboratory

Equipment Location: LAB

Notes: CALIBRATED AT CUSTOMERS SPECIFIED POINTS OF USE ONLY!

| Nominal | Actual (STD) | Measured (UUT) | Deviation (UUT) | Units | Tolerance (+) | Uncertainty (±) | Pass/Fail |
|---------|--------------|----------------|-----------------|-------|---------------|-----------------|-----------|
| 0       | 0.028        | 0.08           | 0.05            | °C    | 0.20          | 0.09            | PASS      |
| 20      | 20.017       | 20.15          | 0.13            | °C    | 0.22          | 0.09            | PASS      |
| 35      | 35.003       | 35.20          | 0.20            | °C    | 0.24          | 0.09            | PASS      |

Deviation rounded to the readability of UUT

The measurement traceability and calibration process used for conformance verification of the above instrument meets or exceeds the requirements of 17025:2005. The reported uncertainties reflect those of type B (Systematic errors associated with the standards and the procedure used), and type A (Random errors of the process). The type A and type B uncertainties were calculated in accordance with NIST technical Note 1297 using the RSS method and are reported at the coverage factor  $k=2$  to approximate a confidence level of 95%. The due date as it appears on this report does not imply that the instrument will maintain its accuracy for any given length of time unless supported with further documentation (E.g. statistical etc.) which affirms such stability and is the responsibility of the end user. Many factors may contribute to instrument in-accuracy over time such as drift, environment, transportation, frequency of use etc. The reported results reflect readings obtained at the time of test only. The reported uncertainties reflect those associated with the calibration process itself and not the instrument under test. If the UUT is a digital electronic measurement instrument add 0.5 of the least significant digit to the above stated uncertainty. The instrument is considered to be in-tolerance based on the observed results (Deviation or departure from nominal value) falling anywhere within its specified tolerance limits without consideration of applied uncertainty. This document shall not be reproduced except in full without the written approval of Q.C. Services, Inc. Procedure Used QCS 3015 (ORIG) (QCSTD 030106-3)

### TRACEABLE STANDARDS USED:

|                  |             |                  |   |
|------------------|-------------|------------------|---|
| Fluke 1522       | S/N: A6C265 | Cal Due: 10/2012 |   |
| ERTCO-EUTECHNICS | S/N: 304526 | Cal Due: 01/2013 | X |
| HART SCI 1502    | S/N 8B552   | Cal Due: 04/2013 | X |

Certified by: Howard Richard

Date: 09/26/2012

Approved By:

Title: Metrologist

Date: 09/26/2012

Example

1

**Good Certificate**

Meets all items  
under "A" from  
the Checklist

Example

2

# CERTIFICATE OF CALIBRATION AND TEST

REF ILR245 SN 2450  
Date 12/25/2012

This product was assembled, tested and calibrated in accordance with the product specifications and FDA Quality System Regulations prior to release for shipment on the date indicated above. Product utilizes calibrated instrumentation traceable to NIST standards in the design, manufacturing, and inspection processes. The calibration results for this products chamber temperature monitoring system are recorded below.

|  |                |  |                |
|--|----------------|--|----------------|
| NIST Factory Thermometer Reading: <u>22</u> °C | ID# <u>010</u> | NIST Factory Thermometer Reading (Lower): (if applicable) <u>22</u> °C | ID# <u>010</u> |
| Product Monitor Probe Reading: <u>22</u> °C    |                | Product Monitor Probe Reading (Lower): (if applicable) <u>22</u> °C    |                |



SIGNATURE \_\_\_\_\_ DATE 1/2/2013

**Example  
2**

# CERTIFICATE OF CALIBRATION AND TEST

REF ILR245 SN 2450  
Date 12/25/2012

This product was assembled, tested and calibrated in accordance with the product specifications and FDA Quality System Regulations prior to release for shipment on the date indicated above. Product utilizes calibrated instrumentation traceable to NIST standards in the design, manufacturing, and inspection processes. The calibration results for this products chamber temperature monitoring system are recorded below.

|  |                |   |                 |
|--|----------------|---|-----------------|
| NIST Factory Thermometer Reading: <u>22</u> °C | ID# <u>010</u> | NIST Factory Thermometer Reading <u>22</u> °C       | ID# <u>010</u>  |
| Probe <u>22</u> °C                             |                | (Lower): (if applicable)                            |                 |
|  |                | Product Monitor Probe Reading (Lower): <u>22</u> °C | (if applicable) |

**Incomplete  
Certificate  
Missing Multiple  
required Items  
from Checklist**



1/2/2013

SIGNATURE DATE

## Report of Validation Primary Temperature Lab

The PRT was calibrated at the following temperatures with the associated uncertainties. The uncertainty evaluation accounts for all known uncertainties present at the time of calibration including long-term behavior of the calibration system, measurement noise, and any short-term effects of the PRT. The uncertainties are reported at the calibration temperatures only. The uncertainties at intermediate temperatures can be computed from these values and the ITS-90 propagation of error curves. The uncertainties are reported at a coverage factor of 2 (k=2).

| CALIBRATION POINT |        |      | TEMPERATURE | MEASURED   | UNCERTAINTY |       |
|-------------------|--------|------|-------------|------------|-------------|-------|
| (point °C)        | (type) | (SN) | t90(°C)     | RESISTANCE | (mK)        |       |
| -197.000          | Comp   | N/A  | -197.000    | 4.6550     | ±6.0        |       |
| -80.000           | Comp   | N/A  | -80.000     | 17.2473    | ±10.0       |       |
| -38.834           | Comp   | N/A  | -38.834     | 21.5122    | ±6.0        |       |
| 0.010             | Comp   | N/A  | 0.010       | 25.4843    | ±4.0        |       |
|                   | In     | FP   | 44013       | 156.599    | 41.0245     | ±6.0  |
|                   | Sn     | FP   | S7005       | 231.928    | 48.2361     | ±6.0  |
|                   | Zn     | FP   | S9007       | 419.527    | 65.4660     | ±9.0  |
|                   | Al     | FP   | 17069       | 660.323    | 86.0321     | ±14.0 |

The following tables indicate the "As Found" RTPW nominal current, the dRTPW in mK, and dRTPW limit in mK. The dRTPW is the change in RTPW during the calibration, not the difference between the "As Found" and "As Left" RTPW. The value of current used in the calibration was 1.000 mA.

| As Found Rtpw | dRtpw Observed | dRtpw Limit |
|---------------|----------------|-------------|
| 1 mA 25.4848  | 0 mK           | 3 mK        |

The following values were determined for the RTPW and the coefficients of the pertinent deviation functions of the ITS-90. For best results, the RTPW value shown should be used as a baseline value for determining the stability of the PRT. The user should maintain a record of RTPW values measured as a routine operation and use these values when computing temperature.

| Model:      | Results for Nominal Current Calibration |
|-------------|---|
| 5628        | RTPW = 25.4843                          |
| Serial No.  | a4 = 3.478321 E-05                      |
| 1819        | b4 = 4.228464 E-06                      |
| Report date | a7 = -2.581569 E-05                     |
| 1/25/13     | b7 = 1.838235 E-05                      |
|             | c7 = -1.226871 E-05                     |

The attached interpolation table was generated from the coefficients listed above. The table is given in terms of resistance (Rt90) versus temperature (°C) at the nominal current. These tables can be used in cases where the readout instrument does not have the capability of computing temperature directly from the coefficients or as a check that the coefficients have been entered into the readout or computer program correctly. The following steps are used to compute temperature from measured resistances utilizing the table. (1) Determine the resistance at the temperature in question. (2) On the table, locate the two resistance values which surround the measured resistance. (3) Subtract the lower of the two from the measured resistance. (4) Divide the result by the sensitivity (dR/dt) from the adjacent column. (5) Add the product of this computation to the temperature which corresponds to the resistance value used in step (3). The additional uncertainty in the tabulated values is negligible (<=0.01mK) but when these tables are used, an additional uncertainty of approximately 0.1 mK should be assumed as a result of the required linear interpolation operation outlined above.

| Nominal | Actual      | Measured    | Error       | Calibration Tolerance | Pass/Fail |
|---------|-------------|-------------|-------------|-----------------------|-----------|
| 0.25    | 0.249996678 | 0.249996716 | 0.000000038 | ±0.000000250          | P         |
| 1.0     | 0.9999107   | 0.9999104   | -0.0000003  | ±0.0000010            | P         |
| 4.0     | 3.9997406   | 3.9997418   | 0.0000012   | ±0.0000040            | P         |

This calibration is traceable to NIST and calibration is compliant to NCSL/ISO/IEC 17025:2005.



Performed by:   
Mike Mike  
Calibration Manager

# Report of Validation Primary Temperature Lab

The PRT was calibrated at the following temperatures with the associated uncertainties. The uncertainty evaluation accounts for all known uncertainties present at the time of calibration including long-term behavior of the calibration system, measurement noise, and any short-term effects of the PRT. The uncertainties are reported at the calibration temperatures only. The uncertainties at intermediate temperatures can be computed from these values and the ITS-90 propagation of error curves. The uncertainties are reported at a coverage factor of 2 (k=2).

| CALIBRATION POINT |        |          | TEMPERATURE | MEASURED   | UNCERTAINTY<br>(mK) |
|-------------------|--------|----------|-------------|------------|---------------------|
| (point °C)        | (type) | (SN)     | t90(°C)     | RESISTANCE |                     |
| -197.000          | Comp   | N/A      | -197.000    | 4.6550     | ±6.0                |
| -80.000           | Comp   | N/A      | -80.000     | 17.2473    | ±10.0               |
| -38.834           | Comp   | N/A      | -38.834     | 21.5122    | ±6.0                |
| 0.010             | Comp   | N/A      | 0.010       | 25.4843    | ±4.0                |
|                   | In     | FP 44013 | 156.599     | 41.0245    | ±6.0                |
|                   | Sn     | FP S7005 | 231.928     | 48.2361    | ±6.0                |
|                   | Zn     | FP S9007 | 419.527     | 65.4660    | ±9.0                |
|                   | Al     | FP 17069 | 660.323     | 86.0321    | ±14.0               |

The following tables indicate the "As Found" RTPW nominal current, the dRTPW in mK, and dRTPW limit in mK. The dRTPW is the change in RTPW during the calibration, not the difference between the "As Found" and "As Left" RTPW. The value of current used in the calibration was 1.000 mA.

| As Found Rtpw | dRtpw Observed | dRtpw Limit |
|---------------|----------------|-------------|
| 1 mA 25.4848  | 0 mK           | 3 mK        |

The following values were determined for the RTPW and the coefficients of the pertinent deviation functions of ITS-90. For best results, the RTPW value shown should be used as a baseline value for determining the PRT. The user should maintain a record of RTPW values measured as a routine operation and use a computer to compute temperature.

Model:  
**5628**

Serial No.  
**1819**

Report date  
**1/25/13**

### Results for Nominal Current Calibration

RTPW = 25.4843  
a4 = 3.478321 E-05  
b4 = 4.228464 E-06  
a7 = -2.581569 E-05  
b7 = 1.838235 E-05  
c7 = -1.226871 E-05

**Good Certificate  
Meets all required  
items under "B"  
from the Checklist**

The attached interpolation table was generated from the coefficients listed above. The table is given in terms of resistance (Rt90) versus temperature (°C) at the nominal current. These tables can be used in cases where the readout instrument does not have the capability of computing temperature directly from the coefficients or as a check that the coefficients have been entered into the readout or computer program correctly. The following steps are used to compute temperature from measured resistances utilizing the table. (1) Determine the resistance at the temperature in question. (2) On the table, locate the two resistance values which surround the measured resistance. (3) Subtract the lower of the two from the measured resistance. (4) Divide the result by the sensitivity (dR/dt) from the adjacent column. (5) Add the product of this computation to the temperature which corresponds to the resistance value used in step (3). The additional uncertainty in the tabulated values is negligible (<=0.01mK) but when these tables are used, an additional uncertainty of approximately 0.1 mK should be assumed as a result of the required linear interpolation operation outlined above.

| Nominal | Actual      | Measured    | Error       | Calibration Tolerance | Pass/Fail |
|---------|-------------|-------------|-------------|-----------------------|-----------|
| 0.25    | 0.249996678 | 0.249996716 | 0.000000038 | ±0.000000250          | P         |
| 1.0     | 0.9999107   | 0.9999104   | -0.0000003  | ±0.0000010            | P         |
| 4.0     | 3.9997406   | 3.9997418   | 0.0000012   | ±0.0000040            | P         |

This calibration is traceable to NIST and calibration is compliant to NCSL/ISO/IEC 17025:2005.

**Example  
3**

Performed by:



Mike Mike  
Calibration Manager

**Calibration complies with ISO/IEC  
17025, ANSI/NCSL Z540-1, and 9001**

Cert. No.: 404

**Certificate of Calibration for Monitoring Thermometer**

Cust ID: Dept Public Hlth.  
RMA:972198 )

**Instrument Identification:**

Model: 61161-2                      S/N: 1116649                      Manufacturer: ConCon

**Standards/Equipment:**

| Description                         | Serial Number | Due Date | NIST Traceable Reference |
|-------------------------------------|---------------|----------|--------------------------|
| Temperature Calibration Bath TC-231 | A79341        |          |                          |
| Thermistor Module                   | A17118        | 2/01/13  | 1000311439               |
| Temperature Probe                   | 3039          | 2/14/13  | 6-BN9WZ-1-1              |
| Temperature Calibration Bath TC-275 | A9A237        |          |                          |
| Digital Thermometer                 | 122044330     | 1/24/13  | 4000-4146811             |

**Certificate Information:**

Technician: 6                      Procedure: CAL                      Cal Date: 9/06/12                      Cal Due: 9/06/17  
Test Conditions: 26.5°C    38.0 %RH    1012 mBar

**Calibration Data:**


| Unit(s)  | Nominal | As Found | In Tol | Nominal | As Left | In Tol | Min  | Max  | ±U   | TUR  |
|----------|---------|----------|--------|---------|---------|--------|------|------|------|------|
| °C Probe |         | N.A.     |        | 0.00    | 0.6     | Y      | -1.0 | 1.0  | 0.06 | >4:1 |
| °C Probe |         | N.A.     |        | 25.00   | 25.5    | Y      | 24.0 | 26.0 | 0.06 | >4:1 |

**This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.**

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. If tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=(Max-Min)/2; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

This certificate indicates calibration for external sensor only.

  
\_\_\_\_\_  
Rodi Wallace, Quality

  
\_\_\_\_\_  
Berry Nic, Technical

**Maintaining Accuracy:**

In our opinion once calibrated your Monitoring Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Monitoring Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

**Recalibration:**

For factory calibration and re-certification traceable to National Institute of Standards and Technology, contact Control Company.

**Example**  
**4**



Calibration complies with ISO/IEC  
17025, ANSI/NCSL Z540-1, and 9001

Cert. No.: 404

Certificate of Calibration for Monitoring Thermometer

Cust ID: Dept Public Hlth,  
RMA:972198 )

Instrument Identification:

Model: 61161-2

S/N: 1116649

Manufacturer: ConCon

Standards/Equipment:

| Description                         | Serial Number | Due Date | NIST Traceable Reference |
|-------------------------------------|---------------|----------|--------------------------|
| Temperature Calibration Bath TC-231 | A79341        |          |                          |
| Thermistor Module                   | A17118        | 2/01/13  | 1000311439               |
| Temperature Probe                   | 3039          | 2/14/13  | 6-BN9WZ-1-1              |
| Temperature Calibration Bath TC-275 | A9A237        |          |                          |
| Digital Thermometer                 | 122044330     | 1/24/13  | 4000-4146811             |

Certificate Information:

Technician: 6 Procedure: CAL Cal Date: 9/06/12 Cal Due: 9/06/17  
Test Conditions: 26.5°C 38.0°C

Results

Uncertainty

Calibration Data:

| Unit(s)  | Nominal | As Found | In Tol | Nominal | As Left | In Tol | Min  | Max  | ±U   | TUR  |
|----------|---------|----------|--------|---------|---------|--------|------|------|------|------|
| °C Probe |         | N.A.     |        | 0.00    | 0.6     | Y      | -1.0 | 1.0  | 0.06 | >4:1 |
| °C Probe |         | N.A.     |        | 25.00   | 25.5    | Y      | 24.0 | 26.0 | 0.06 | >4:1 |

Pass/Fail or  
In Tolerance

This instrument was calibrated using instruments traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty of the instrument used for calibration. The uncertainty of the instrument used for calibration is determined by using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction in tolerance. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=(Max-Min)/2; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

This certificate indicates calibration for external service only.

Rod Wallace, Quality

Berry Nic, Technical

Maintaining Accuracy:

In our opinion once calibrated your Monitoring Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Monitoring Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contaminants.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Example

4

Good Certificate  
Meets all required  
items under "B"  
from the Checklist

# REPO CALIBRATION REPORT

Certificate # 130

Instrument ID 162      Description  
 Manufacturer CAR      Model Number 5000T  
 Calibrated 3/19/2013      Frequency Annual  
 Serial Number 7620023  
 Next Cal Date 3/19/2014

| <u>Calibration Specifications</u> |         |              |                   |     |        |       |        |           |  |
|-----------------------------------|---------|--------------|-------------------|-----|--------|-------|--------|-----------|--|
| Group # 1                         |         |              |                   |     |        |       |        |           |  |
| Group Name 2.PT.CAL               |         |              |                   |     |        |       |        |           |  |
| Nom In Val / In Val               | In Type | Std Accy     | Acc %             | ±L  | Ind As | Lt As | Dev %  | Pass/Fail |  |
| 5.0 / 5.0                         | C       | Plus / Minus | 0.00000 / 0.00000 | 0.5 | 5.0    | 5.0   | 0.00%  | Pass      |  |
| -15.0 / -15.0                     | C       | Plus / Minus | 0.00000 / 0.00000 | 0.5 | -15.0  | -14.5 | -3.33% | Pass      |  |

Test Instruments Used During the Calibration

| Test Instrument ID    | Description | Manufacturer    | Model Number | Serial Number | (As Of Cal Entry Date)                             |
|-----------------------|-------------|-----------------|--------------|---------------|--|
| HART PRECISION<br>RTD |             | HART-SCIENTIFIC | 1502A        | A1B599        | Last Cal Date 5/21/2012<br>Next Cal Date 5/21/2013 |

Notes about this calibration

Company Inc. certifies that the above equipment has been calibrated using instrumentation and standards that are traceable to the National Institute of Standards and Technology (NIST) through certification documents on file. This calibration complies with MIL-STD-45662A and ISO 17025, Test Uncertainty Ratio ≥ 4:1 unless otherwise stated.

**Example  
5**

Calibration Result    Calibration Successful  
 Who Calibrated    Davis Calvin

Finalized By    Huson  
 Date Finalized    3/19/2013

# REPO CALIBRATION REPORT

Certificate # 130

Instrument ID 162

Manufacturer CAR

Calibrated 3/19/2013

Description

Model Number 5000T

Frequency Annual

Serial Number 7620023

Next Cal Date 3/19/2014

| Group # 1           |         | Group Name 2.FT.CAL |              | Calibration Specifications |     | Results  |        |       |        |           |
|---------------------|---------|---------------------|--------------|----------------------------|-----|----------|--------|-------|--------|-----------|
| Nom In Val / In Val | In Type | Std Accy            | Plus / Minus | Acc %                      | ±L  | Out Type | Inp As | Lt As | Dev %  | Pass/Fail |
| 5.0 / 5.0           | C       | 0.000000 / 0.000000 | Plus / Minus | 0.000000 / 0.000000        | 0.5 | C        | 5.0    | 5.0   | 0.00%  | Pass      |
| -15.0 / -15.0       | C       | 0.000000 / 0.000000 | Plus / Minus | 0.000000 / 0.000000        | 0.5 | C        | -14.5  | -14.5 | -3.33% | Pass      |

| Test Instruments Used During the Calibration |             | Manufacturer    |       | Model Number |           | Serial Number |           | (As Of Cal Entry Date) |               |
|--|-------------|-----------------|-------|--------------|-----------|---------------|-----------|------------------------|---------------|
| Test Instrument ID                           | Description | HART-SCIENTIFIC | 1302A | A1B599       | 5/21/2012 | 5/21/2012     | 5/21/2012 | Next Cal Date          | Next Cal Date |
| HART PRECISION                               | RTD         |                 |       |              |           |               |           |                        |               |

Notes about this calibration

Company Inc. certifies that the above equipment has been calibrated using instrumentation and standards that are traceable to the National Institute of Standards and Technology (NIST) through certification documents on file. This calibration complies with MIL-STD-45662A and ISO 17025, Test Uncertainty Ratio ≥ 4:1 unless otherwise stated.

Example

5

**Good Certificate**

**Meets all required items under "B" from the Checklist**

Calibration Result Calibration Successful  
Who Calibrated Davis Calvin

Finalized By Huson  
Date Finalized 3/19/2013

# INSTRUMENT CALIBRATION REPORT

CDC

Certificate # 4701

Instrument ID 16238  
 Manufacturer LASCAR  
 Calibrated 3/19/2013

Description  
 Model Number VJ5000T  
 Frequency Annual

Serial Number 010023762  
 Next Cal Date 3/19/2014


| Nom In Val / In Val |  | In Type |  | Std Accy     |  | Acc %               |  | ±/  |  | End As |  | Out Type |  | Lft As |  | Dev %  |  | Pass/Fail |  |
|---------------------|--|---------|--|--------------|--|---------------------|--|-----|--|--------|--|----------|--|--------|--|--------|--|-----------|--|
| 5.0 / 5.0           |  | C       |  | Plus / Minus |  | 0.000000 / 0.000000 |  | 0.5 |  | 5.0    |  | C        |  | 5.0    |  | 0.00%  |  | Pass      |  |
| -15.0 / -15.0       |  | C       |  | Plus / Minus |  | 0.000000 / 0.000000 |  | 0.5 |  | -15.0  |  | C        |  | -14.5  |  | -3.33% |  | Pass      |  |

| Test Instruments Used During the Calibration |  |
|--|--|
| Test Instrument ID<br>HART PRECISION<br>RTD  | Description<br>HART SCIENTIFIC<br>1502A                                      |
| Manufacturer<br>HART SCIENTIFIC              | Model Number<br>1502A  |
| Serial Number<br>AIB599                      | (As Of Cal Entry Date)<br>Last Cal Date 5/21/2012<br>Next Cal Date 5/21/2013 |

Notes about this calibration

SolConut certifies that the above equipment has been calibrated using instrumentation and standards that are traceable to the National Institute of Standards and Technology (NIST) through certification documents on file. This calibration complies with MIL-STD-45662A and ISO 10012-1 and ANSI/NCSL Z540-1-1994. Test Uncertainty Ratio  $\geq 4:1$  unless otherwise stated.



**SolConut**  
 Phone: (888) 555-0636  
 Fax: (555) 555-5419

Calibration Result Calibration Successful  
 Who Calibrated Cble Hu

Finalized By Crav Swin  
 Date Finalized 3/19/2013 10:52:24AM

# INSTRUMENT CALIBRATION REPORT

Certificate # 4701

CDC

|                      |                         |
|----------------------|-------------------------|
| Instrument ID 16238  | Description             |
| Manufacturer LASCAR  | Model Number VJ5000T    |
| Calibrated 3/19/2013 | Frequency Annual        |
|                      | Serial Number 010023762 |
|                      | Next Cal Date 3/19/2014 |

| <u>Calibration Specifications</u> |   |              |   |                     |       |       |          |        |        |        |           |
|-----------------------------------|---|--------------|---|---------------------|-------|-------|----------|--------|--------|--------|-----------|
| Nom In Val / In Val               |   | In Type      |   | Std Accy            | Acc % | ±L    | Out Type | End As | Lft As | Dev %  | Pass/Fail |
| 5.0 / 5.0                         | C | Plus / Minus | C | 0.000000 / 0.000000 | 0.5   | 5.0   | C        | 5.0    | 5.0    | 0.00%  | Pass      |
| -15.0 / -15.0                     | C | Plus / Minus | C | 0.000000 / 0.000000 | 0.5   | -15.0 | C        | -14.5  | -14.5  | -3.33% | Pass      |

Test Instruments Used During the Calibration

|                       |             |                 |              |               |               |               |
|-----------------------|-------------|-----------------|--------------|---------------|---------------|---------------|
| Test Instrument ID    | Description | Manufacturer    | Model Number | Serial Number | Last Cal Date | Next Cal Date |
| HART PRECISION<br>RTD |             | HART SCIENTIFIC | 1502A        | AJB599        | 5/21/2012     | 5/21/2013     |

(As Of Cal Entry Date)

Notes about this calibration

Solconut certifies that the above equipment has been calibrated using instrumentation and standards that are traceable to the National Institute of Standards and Technology (NIST) through certification documents on file. This calibration complies with MIL-STD-45662A and ISO 10012-1 and ANSI/NCSL Z540-1-1994. Test Uncertainty Ratio  $\geq$  4:1 unless otherwise stated.

**Example**  
6

**Solconut**  
Phone: (888) 555-0636  
Fax: (555) 555-5419

**Incomplete  
Certificate  
Missing  
ISO 17025  
Statement**

Calibration Rest  
Who Calibrat

Finalized By Crav Swin  
Date Finalized 3/19/2013 10:52:24AM



Company  
Corporation  
Street  
USA

Certificate Number: 01845      Ambient Temperature(°C): Min: 21.4    Max: 23.8  
 Model Number: VL-200      Ambient Humidity(%RH): Min: 30.0    Max: 59.6  
 Serial Number: 120521      Method: Calibration by comparison  
 Procedures: VCP1009 VCP1010

The calibration(s) on this report are traceable to the United States of America National Institute of Standards and Technology or to other recognized national or international standards or to accepted values of natural physical constants, and are accredited to ISO/IEC 17025. The laboratory meets the requirements of ANSI/NCSL Z540-1. Using methods detailed in the ISO "Guide to the Expression of Uncertainty in Measurement", reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2. The calibrations were performed equivalently either with minimum test uncertainty ratios of 4:1 using a coverage factor of k = 2, or with the statistical method of guard banding to reduce test limits. The results relate only to the item(s) calibrated.

### CALIBRATION REFERENCE EQUIPMENT



|  | Serial # | Calibration Date |           |
|--|----------|------------------|-----------|
|  |          | Last             | Next      |
| Hart Scientific Black Stack Thermistor Scanner Module Model 2564 | A39287   | 27-Nov-11        | 27-Nov-12 |
| Thunder Scientific Humidity Generator 2500 ST-I.T                | 1007799  | 28-Jul-11        | 28-Jul-12 |
| Hart Scientific Thermistor Temperature Probe Model 5610          | B0B1519  | 18-Jan-12        | 18-Jan-13 |

### CALIBRATION TEST RESULTS

| Chan | Test Description          | Units | Reference | Measurement | As Left |      |
|------|---------------------------|-------|-----------|-------------|---------|------|
|      |                           |       |           | Uncertainty | Result  | Diff |
| 1    | Temperature               | °C    | -25.34    | 0.06        | -25.33  | 0.01 |
| 1    | Temperature               | °C    | 9.64      | 0.05        | 9.65    | 0.01 |
| 1    | Temperature               | °C    | 25.01     | 0.04        | 25.01   | 0.00 |
| 1    | Temperature               | °C    | 44.73     | 0.06        | 44.71   | 0.02 |
| 1    | Temperature               | °C    | 69.55     | 0.07        | 69.55   | 0.00 |
| 2    | Relative Humidity at 10°C | %RH   | 45.00     | 0.60        | 45.52   | 0.52 |
| 2    | Relative Humidity at 25°C | %RH   | 11.00     | 0.60        | 11.34   | 0.34 |
| 2    | Relative Humidity at 25°C | %RH   | 45.00     | 0.60        | 45.26   | 0.26 |
| 2    | Relative Humidity at 25°C | %RH   | 80.00     | 0.60        | 80.27   | 0.27 |
| 2    | Relative Humidity at 45°C | %RH   | 45.00     | 0.60        | 45.27   | 0.27 |

### Maintaining Calibration

The electronic components in this data logger are of the highest quality. The unit has been designed to remain within its specifications. The length of in-calibration service can be affected by aging, temperature and shock. For those users with critical needs such as accreditation demands, government specifications, or ISO requirements, we recommend that the unit be calibrated on a periodic basis.

  
 Calibration Technician: *Crend P*  
  
 Technician: *Paul C*

### Calibration

Information on validation services is available at the address below. This data logger was calibrated by:

Sal Inc.  
 100-Pkwy.  
 Richmond, CA 94874  
 Toll Free: 1-800-555-8374, Phone: 555-555-5850, Fax: 555-555-2874  
 Email: support@sal.com

Calibration Date: 19-Jul-2012

Next Calibration: 19-Jul-2013

**Example**  
7



Company

Corporation  
Street  
USA

Certificate Number: 01845      Ambient Temperature(°C): Min: 21.4    Max: 23.8  
Model Number: VL-200      Ambient Humidity(%RH): Min: 30.0    Max: 59.6  
Serial Number: 120521      Method: Calibration by comparison  
Procedures: VCP1009 VCP1010

The calibration(s) on this report are traceable to the United States of America National Institute of Standards and Technology or to other recognized national or international standards or to accepted values of natural physical constants, and are accredited to ISO/IEC 17025. The laboratory meets the requirements of ANSI/NCSL Z540-1. Using methods detailed in the ISO "Guide to the Expression of Uncertainty in Measurement", reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2. The calibrations were performed equivalently either with minimum test uncertainty ratios of 4:1 using a coverage factor of k = 2, or with the statistical method of guard banding to reduce test limits. The results relate only to the item(s) calibrated.

CALIBRATION REFERENCE EQUIPMENT

|  | Serial # | Calibration Date |           |
|--|----------|------------------|-----------|
|  |          | Last             | Next      |
| Hart Scientific Black Stack Thermistor Scanner Module Model 2564 | A39287   | 27-Nov-11        | 27-Nov-12 |
| Thunder Scientific Humidity Generator 2500 ST-LT                 | 1007799  | 28-Jul-11        | 28-Jul-12 |
| Hart Scientific Thermistor Temperature Probe Model 5610          | B0B1519  | 18-Jan-12        | 18-Jan-13 |

CALIBRATION TEST RESULTS

| Chan | Test Description          | Units | Reference | Measurement | As Left |       |
|------|---------------------------|-------|-----------|-------------|---------|-------|
|      |                           |       |           | Uncertainty | Result  | Diff. |
| 1    | Temperature               | °C    | -25.34    | 0.06        | -25.33  | 0.01  |
| 1    | Temperature               | °C    | 9.64      | 0.05        | 9.65    | 0.01  |
| 1    | Temperature               | °C    | 25.01     | 0.04        | 25.01   | 0.00  |
| 1    | Temperature               | °C    | 44.73     | 0.06        | 44.71   | 0.02  |
| 1    | Temperature               | °C    | 69.55     | 0.07        | 69.55   | 0.00  |
| 2    | Relative Humidity at 10°C | %RH   | 45.00     | 0.60        | 45.52   | 0.52  |
| 2    | Relative Humidity at 25°C | %RH   | 11.00     | 0.60        | 11.34   | 0.34  |
| 2    | Relative Humidity at 25°C | %RH   | 45.00     | 0.60        | 45.26   | 0.26  |
| 2    | Relative Humidity at 25°C | %RH   | 80.00     | 0.60        | 80.27   | 0.27  |
| 2    | Relative Humidity at 45°C | %RH   | 45.00     | 0.60        | 45.27   | 0.27  |



Maintaining Calibration

The electronic components in this data logger are of the highest quality. The unit has been designed to remain within its specifications. The length of in-calibration service can be affected by aging, temperature and shock. For those users with critical needs such as accreditation demands, government specifications, or ISO requirements, we recommend that the unit be calibrated on a periodic basis.

\_\_\_\_\_  
 Calibration Technician: Crend P

\_\_\_\_\_  
 Technician: Nuz Cre

Calibration

Information on calibration services is available at the address below. This data logger was calibrated by:

Sal Inc.  
 100-Pkwy.  
 Richmond, CA 2874  
 Toll Free: 1-800-555-8374, Phone: 555-555-5850, Fax: 555-555-2874  
 Email: support@sal.com,

Calibration Date: 19-Jul-2012

Next Calibration: 19-Jul-2013



|                             |                        |                               |                        |
|-----------------------------|------------------------|-------------------------------|------------------------|
| <b>System ID:</b>           | Aurora Specialty       | <b>Calibration Date/Time:</b> | 2013/04/16 07:35 AM    |
| <b>Component ID:</b>        | RF1-TE2                | <b>Calibrated By:</b>         | Pody Wall              |
| <b>Equipment Type:</b>      | Temperature Loop       | <b>Quality Approval:</b>      | Johnn Loo 2013/04/18   |
| <b>Description:</b>         | TE-02 Temperature Loop | <b>Next Event Due Date:</b>   | 2014/04/30             |
| <b>Instrument Location:</b> | Zone East              | <b>Next Event Name:</b>       | Annual Calibration     |
| <b>Manufacturer:</b>        | Precon                 | <b>Customer Name/Contact:</b> | Phel                   |
| <b>Model:</b>               | ST-S81E                | <b>Customer Location:</b>     | 34 Fraser ST, CA 11800 |
| <b>Serial Number:</b>       | NI800                  | <b>Calibration Results:</b>   | Pass                   |
| <b>Instrument Range:</b>    | -200 to 800 °C         | <b>Out of Tolerance:</b>      | Init.                  |
| <b>Tolerance:</b>           | ± 0.5 °C               | <b>Ambient Condition:</b>     | 69 °F / 27 %RH         |

**CALIBRATION DATA**

|   |        | AS FOUND  |       | AS LEFT    |           |                          |            |
|---|--------|-----------|-------|------------|-----------|--------------------------|------------|
| Parameter:  | Units: | Setpoint: | Data: | Deviation: | Setpoint: | Data:                    | Deviation: |
| NA  | °C     | 15.91     | 15.8  | -0.11      | NP        | NP                       |            |
| NA  | °C     | 1.62      | 1.7   | 0.08       | NP        | NP                       |            |
| NA  | °C     | 5.05      | 5.1   | 0.05       | NP        | NP                       |            |
| <b>Reference Standard Used (Mfg, Model #, Serial # or Lot #, and Calibration Due Date):</b><br>Hart Scientific, 1521, A22097, 2013/07/26<br>Hart Scientific, 5613, 711917, 2013/04/25 |        |           |       |            |           |                          |            |
|   |        |           |       |            |           | <b>Procedure Used:</b>   |            |
|   |        |           |       |            |           | SOP-4-146-05, 2011/12/16 |            |

**Example**  
8

**Comments:**

Reference Standard was placed as close to the U.U.T. as possible to obtain a stable reading.  
 Calibration offset "As Found" 0.0. Calibration offset "As Left" 0.0. KLP 16APR2013



|                      |                        |                        |                        |
|----------------------|------------------------|------------------------|------------------------|
| System ID:           | Aurora Specialty       | Calibration Date/Time: | 2013/04/16 07:35 AM    |
| Component ID:        | RF1-TEZ                | Calibrated By:         | Pody Wall              |
| Equipment Type:      | Temperature Loop       | Quality Approval:      | Johnn Loo 2013/04/18   |
| Description:         | TE-02 Temperature Loop | Next Event Due Date:   | 2014/04/30             |
| Instrument Location: | Zone East              | Next Event Name:       | Annual Calibration     |
| Manufacturer:        | Precon                 | Customer Name/Contact: | PheI                   |
| Model:               | ST-S81E                | Customer Location:     | 34 Fraser ST, CA 11800 |
| Serial Number:       | N1800                  | Calibration Results:   | Pass                   |
| Instrument Range:    | -200 to 800 °C         | Out of Tolerance:      | Init.                  |
| Tolerance:           | ± 0.5 °C               | Ambient Condition:     | 69 °F / 27 %RH         |

**CALIBRATION DATA**

| Parameter: | Units: | AS FOUND  |       |            | AS LEFT   |       |            |
|------------|--------|-----------|-------|------------|-----------|-------|------------|
|            |        | Setpoint: | Data: | Deviation: | Setpoint: | Data: | Deviation: |
| NA         | °C     | 15.91     | 15.8  | -0.11      | NP        | NP    |            |
| NA         | °C     | 1.62      | 1.7   | 0.08       | NP        | NP    |            |
| NA         | °C     | 5.05      | 5.1   | 0.05       | NP        | NP    |            |



|  |                           |
|--|---------------------------|
| Reference Standard Used (Mfg, Model #, Serial # or Lot #, and Calibration Due Date): | Procedure Used:           |
| Hart Scientific, 1521, A22097, 2013/07/26  | SOP-4-146- 05, 2011/12/16 |
| Hart Scientific, 5613, 711917, 2013/04/25  |                           |

**Example**  
8

Comments:  
Reference Standard was placed as close to the U.U.T. as possible to obtain a stable reading.  
Calibration offset "As Found" 0.0. Calibration offset "As Left" 0.0. KLP 16APR2013

Customer Approval (Optional)-Reviewed By/Date: \_\_\_\_\_ Quality Approval/Date: \_\_\_\_\_